




TECHNICAL ENVIRONMENTAL SERVICES, LLC



Mr. Jamie Foster
Department of Environmental Protection
Division of Solid Waste Services
16010 Frederick Road
Rockville, MD 20855

October 13, 2020

Dear Jamie,

Enclosed is your semi-annual calibration report for the meteorological calibration performed on October 6, 2020 at the East Gude Drive Landfill site. All instrumentation was compared against NIST certified test equipment and found to be functioning correctly. The rain gauge needs to be watched closely since spiders like it as a location for spinning webs.


Please let me know if you have any questions about the calibration.

Best regards,

Mark Abrams
President - TES



TECHNICAL ENVIRONMENTAL SERVICES, INC.
P.O. BOX 3253
GAITHERSBURG, MD 20885



MIDAS SYSTEM CALIBRATION RESULTS

FOR: MONTGOMERY COUNTY, MD (EAST GUDE DR)

LOCATION: ROCKVILLE, MD

LAT: 39°6.47'N/LONG: 77°8.42'W

MAGNETIC VARIATION: 4.0°W(-)

SITE ELEVATION: 400 FT MSL

TOWER HEIGHT: 3 M (10 ft)

DATE OF CALIBRATION: October 6, 2020

SYSTEM ID: 170630-001

SYSTEM TYPE: MIDAS/CLIMATRONICS/WM283

EQUIPMENT: CLIMATRONICS CORP:

101283	WIND SPEED
101283	WIND DIRECTION
HMP 60	TEMPERATURE
HMP 60	RELATIVE HUMIDITY
TE-525	PRECIPITATION
CR-300	DATA LOGGER; SN 5100

PROCEDURES USED: TECHNICAL ENVIRONMENTAL SERVICES 508 (series)

TEST EQUIPMENT USED FOR CALIBRATION:

<u>TYPE</u>	<u>MODEL</u>	<u>SER NO.</u>
Compass, digital	2074	1000057
Digital voltmeter	2030E	50811075
Torque Disc	18310	None
Thermometer, LIG	FP-10	3N8751
Hygro-Therm (RH/T)	445580	913785
Anemometer Drive	18801	CA 01550
Speedtech Instruments	Skymaster	3396

SYSTEM EQUIVALENTS

THIS IS DIGITAL DATA COLLECTION SYSTEM; THERE ARE NO VOLTAGE EQUIVALENTS

Signature of technician
performing tests:

[Handwritten Signature]

Date: 10/6/20

Reviewed by:

[Handwritten Signature]

Date: 10/6/20

POWER SUPPLY AND MISC TESTS

PROCEDURE: 508.10

Page 1 of 8

BATTERY TESTS:

- 1. UNLOADED VOLTAGE: NA
- 2. CHARGER VOLTAGE: 13.75
- 3. VOLTAGE AT DATA LOGGER: 13.78 (Should be 12.8>)

4. REMARKS: All instrumentation was found to be working correctly.

5. FINDINGS TO BE ADDRESSED BY OWNER: None

DATE OF ALL TESTS: 10/6/20

DATA SHEET - WIND SPEED

PROCEDURE: 508.1

9.0 RECORD OF CALIBRATION

9.1 Serial Numbers: SENSOR: 154832

PROPELLER: 08245

9.2 Torque Values Measured:

AS FOUND

AFTER BEARING/SENSOR REPLACEMENT

1.0 (gm-cm)

NA (gm-cm)

Bearings replaced: Yes No

[Note: 1.0 gm-cm = 1 mph; 3.8 gm-cm = 2.0 mph]

9.3 If new sensor or propeller are required record new serial numbers here:

SENSOR: NA

PROPELLER: NA

IF A COLLOCATED TRANSFER STANDARD METHOD IS USED TO CHECK THE VIABILITY OF THE OPERATIONAL SENSOR RECORD THE RESULTS BELOW.

Expected

CTS READING:

COMPUTER READS:

RPMS:

11.4

1000

5

3000

5000

ALLOWABLE TOLERANCE IS ±0.5 mph

DATA SHEET - WIND DIRECTION

PROCEDURE: 508.2

Page 3 of 8

9.0 RECORD OF CALIBRATION

9.1 Serial Numbers:

SENSOR: 154832

9.2 Torque Values Measured:

AS FOUND

AFTER BEARING/SENSOR REPLACEMENT

6.0 (gm-cm)

NA (gm-cm)

Bearings replaced: Yes No

[Note: 11 gm-cm = 1.0 mph; 25 gm-cm = 2 mph]

9.3 If new sensor is required record new serial number here:

SENSOR: NA

9.4 Linearity test of direction sensor; "as found" or "new", (circle one):

Expected	Degrees (At Computer)
360 deg	<u>357</u>
090 deg	<u>091</u>
180 deg	<u>183</u>
270 deg	<u>271</u>

ALLOWABLE TOLERANCE IS $\pm 5^\circ$

DATA SHEET - WIND DIRECTION ALIGNMENT

PROCEDURE: 508.2.4

Page 4 of 8

9.0 RECORD OF CALIBRATION

9.1 Serial Numbers: 154832

9.2 Reference orientation point: North or South (true) by compass.

Magnetic variation: 4.0°W (-); 4°Mag = 360°True/184° = 180°True

9.3 "As Found" alignment before any adjustments; ALLOWABLE TOLERANCE IS

358.2

9.4 "After Adjustment" alignment (see note):

NA

Note: If any repairs to a sensor are accomplished or if a replacement sensor is installed indicate action(s) here and perform tests on the new or refurbished sensor and complete a second test of alignment and linearity.

DATA SHEET - TEMPERATURE

PROCEDURE: 508.3

Page 5 of 8

9.0 RECORD OF TEMPERATURE VERIFICATION TESTS

9.1 Serial Numbers:

SENSOR: R 3/5/018

9.1 Ambient test of system:

Ambient = 66.8 °F; system read: 66.1 °F

(Allowable tolerance is $\pm 0.9^\circ\text{F}$)

NOTES OR REMARKS AND ANY ADDITIONAL TESTS; RECORD RESULTS HERE:

DATA SHEET - RELATIVE HUMIDITY

PROCEDURE: 508.4

Page 6 of 8

9.0 RECORD OF HUMIDITY COMPARISON TESTS

9.1 Serial Numbers:

SENSOR: R3151018

9.2 Ambient humidity calculation at:

1200 (Time)

Relative Humidity from co-located
transfer standard

53.3 %

System relative humidity from
computer display/keypad:

56.7 %

Allowable tolerance is $\pm 5\%$ between the derived value
and the computer value for RH at temperatures above 32°F.

Note: If any repairs are made or the humidity sensor is replaced indicate
action(s) here and perform tests on new or repaired equipment on a second
datasheet.

DATA SHEET - PRECIPITATION

PROCEDURE: 508.12

Page 8 of 8

9.0 RECORD OF PRECIPITATION GAUGE TESTS

9.1 Serial Numbers:

GAUGE: 72000-217

9.2 Measured amount of water deposited in the catch opening: 5.2 ml

Left

Right

5.2

5.2

ALLOWABLE TOLERANCE $\pm 10\%$ OF OBSERVED OR $\pm 0.5\text{MM}$

Test Equipment Certification Record Technical Enviromental Services, LLC

14-May-20

Page 1of 1

Instrument	Manufacturer	Model No.	Serial No.	Last Cal.	Next Cal.
Anemometer Drive	R.M. Young Co.	18801	CA 01550	5/4/2020	5/4/2021
Digital Compass	Autohelm	2074	810753	Fixed Device	NA
Digital Multimeter	Wavetek	2030E	50811075	1/16/2020	1/16/2022
Psychrometer	Belfort Instruments	566-2 (T3 & 4)	85-608	1/15/2020	1/15/2022
Torque Disc	R.M. Young Co.	18310	None	Fixed Device	NA
Torque Gauge	R.M. Young Co.	18331	None	Fixed Device	NA
Barometer	Baltic	2215	None	Fixed Device	NA
Hygro-Therm Device	Extech Instruments	445580	913785	10/9/2019	10/9/2021
Ambient Weather	Ambient Weather	WM-4	NA	4/23/2019	4/23/2021
Global Radiometer	Solar Light Co.	PMA1140	6475	2/11/2020	2/10/2021